

Certificate of Conformity

 (\bullet)

Number CoC-2380227-01 Project number 2380227 Page 1 of 3

Issued by	: NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands	
Applicant	: Algodue Elettronica S.r.l. Via Piero Gobetti, 16/F 28014 Maggiora (NO) Italy	
Submitted	A meter embedding IEC 61000-4-30 class A Power Quality functions Manufacturer : Algodue Type : PQM3000, PQM3000RGW PQM4000, PQM4000RGW	
Characteristics	: See page 2 and further	
In accordance with	: IEC 61000-4-30 Ed. 3 (2015) "Electromagnetic Compatibility (EMC) – Part 4-30: Testing and measurement techniques – Power quality measurement methods"	
Measurement class	: IEC 61000-4-30 class A	

The undersigned declares that the described product is tested according to the above mentioned standard and meet their requirements, based on a non-recurrent examination. The appertaining test data is presented in type evaluation report number NMi-15200632-01 and NMi-2380227-01, granted by NMi Certin B.V.

NMi Certin B.V. 24 October 2019

C. Oosterman Head Certification Board

NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document only is permitted.







Certificate of Conformity

Number CoC-2380227-01 Project number 2380227 Page 2 of 3

IEC 61000-4-30 Power Quality functions tested

The following IEC 61000-4-30 measurement methods have been tested

+

Table 1 IEC 61000-4-30 Power Quality functions tested

IEC 62586-2 Clause	Parameter	IEC 61000-4-30 class	Comments
6.1	Power frequency	Α	50 and 60 Hz
6.2	Magnitude of supply voltage	Α	
6.3	Flicker	А	Class F3: 230V, 50 Hz
6.4	Supply voltage interruptions, dips and swells	A	
6.5	Supply voltage unbalance	Α	
6.6	Voltage harmonics	А	
6.7	Voltage interharmonics	Α	
6.8	Mains signalling voltages on the voltage supply	А	Method 1 + 2 (dynamic)
6.9	Measurement of underdeviation and overdeviation parameters	А	
6.10	Flagging	Α	
6.11	Clock uncertainty testing	Α	
6.12	Variation of external influence quantities	A	Temperature range: -25°C +55°C AUX Supply type 1: 85 - 265 VAC 65 - 250 VDC AUX Supply type 2: 19 - 60 VDC
6.13	Rapid Voltage Changes (RVC)	A	
6.14	Magnitude of current		
6.15	Harmonic current	Α	
6.16	Interharmonic currents	Α	
6.17	Current unbalance	Α	
8	Calculation of measurement uncertainty and operating uncertainty	A	

A : compliance with class A

S : compliance with class S

--- : Not implemented

The tests are performed in accordance with IEC 62586-2 edition 2 (2017).

The current tests only apply to the direct current inputs of the PQM3000 and PQM4000.



Certificate of Conformity

(+)

Number CoC-2380227-01 Project number 2380227 Page 3 of 3



Characteristics of the measuring instrument

In Table 2 the general characteristics of the measuring instrument are presented.

Table 2 General characteristics

U _{din}	230 V _{LN}		
U _{max}	345 V _{LN}		
I _{nom}	1 A or 5 A		
fnom	50 Hz and 60 Hz		
Temperature	Rated range of operation: -25°C to +55°C		
AUX. Power supply range	Type 1: 85 – 265 VAC 50/60 Hz 65 – 250 VDC		
	Type 2: 19-60 VDC		
Software version	v.1.15.1.18		
Hardware version 🕒	PQM4000 configuration 3.3 PQM3000 configuration 3.3		
Environmental application	Fixed (F), Indoor (I)		